

# NIST

## Calibration Certificate



The instrument identified below was tested in accordance with factory calibration procedures under the conditions noted below using standards traceable to the National Institute of Standards and Technology (NIST).

**Instrument**

Description:

Serial #:

**Measurement Standards**

Description: 4340 Steel Blocks

Serial #s: 92-5617      97-7715  
 04-6524      04-6340  
 93-5716      93-5718  
 95-5571      97-8418

Report #: 04-25714-A & 04-28146-A

**Conditions**

Temperature: 22°C (72°)

Humidity: 62% RH

**Test Results**

**Date:**

| Check Point (Inch) | Allowable Deviation | Actual Reading | Check Point (Inch) | Allowable Deviation | Actual Reading |
|--------------------|---------------------|----------------|--------------------|---------------------|----------------|
| 0.040              | ±0.004              | 0.040          | 0.300              | ±0.004              | 0.300          |
| 0.050              | ±0.004              | 0.052          | 0.400              | ±0.004              | 0.400          |
| 0.060              | ±0.004              | 0.060          | 0.500              | ±0.004              | 0.500          |
| 0.075              | ±0.004              | 0.076          | 0.750              | ±0.004              | 0.752          |
| 0.100              | ±0.004              | 0.100          | 1.000              | ±0.004              | 1.000          |
| 0.200              | ±0.004              | 0.200          | 2.000              | ±0.004              | 2.000          |