



**NIDEC-SHIMPO AMERICA CORPORATION
INSTRUMENT DIVISION**

NIDEC-SHIMPO (ZHEJIANG) CORPORATION
1858 Pingcheng Street, Pinghu Economic
Development Zone, Zhejiang, China
TEL 86-573-85093022 FAX 86-573-85093543

CERTIFICATE OF CALIBRATION

THIS IS TO CERTIFY THAT SHIMPO INSTRUMENT IDENTIFIED BELOW HAS BEEN TESTED AT OUR FACTORY IN ACCORDANCE WITH APPROVED PROCEDURES. THIS IS ALSO TO CERTIFY THAT THE STANDARDS USED IN THE CALIBRATION OF THIS UNIT ARE PERIODICALLY INSPECTED AND ARE TRACEABLE TO BEIJING AEROSPACE INSTITUTE FOR METROLOGY AND MEASUREMENT TECHNOLOGY

MODEL : TNP-2 SERIAL NO.: *P9217X D004*
DATE CERTIFIED: *Oct 15 / 2017* CAPACITY: 2.000N·m x 0.001N·m
ACCURACY: ±0.5% OF FULL SCALE

CALIBRATION WHEEL: TN-E31No.C24001

WEIGHTS USED:

Base A 400g, 0.4-① 400g, 0.4-② 400g, 0.4-③ 400g, 0.4-④ 400g

OFFICIAL INSPECTION: NIDEC-SHIMPO(ZHEJIANG)CORPORATION
1858 Pingcheng Road, Pinghu Economic Development Zone, Zhejiang, China

CALIBRATED by

Jia yue zhu

NIDEC-SHIMPO (ZHEJIANG) CORPORATION