




# Certificate of Calibration

**Certificate Number:** C1157

**Issue Date:** 2018-01-01

**THIS IS TO CERTIFY THAT THE EQUIPMENT DETAILED HERE-ON CONFORMS TO THE PUBLISHED SPECIFICATIONS AND HAS BEEN TESTED AGAINST OUR TEST INSTRUMENT WHICH IS TRACEABLE TO NATIONAL LABORATORY STANDARDS**

**Instrument Type:** MT2013  
**Serial Number:** SAMPLE  
**Manufacturer:** Compact Instruments Ltd  
**Calibration:** +/- 0.01% +/- 1 digit  
**Submitted By:**  
**Operational Check:** Good  
**Physical Check:** Good  
**Authorised Signatory:** 

**Test Equipment Used**  
Off Air Frequency Standard Model 2A-01

**Traceable to**  
UKAS Approved Calibration Laboratories

**Certificate Number**  
U256696 / U256697

## Test Results

Test Rig rpm	Unit Under Test rpm	Deviation
60000	60000	+0.000%
6000	6000	+0.000%
600	600	+0.000%
60	60	+0.000%
6	6	+0.000%

**Compact Instruments Ltd**  
61-65 Lever Street  
Bolton  
Lancashire  
BL3 2AB

**Phone:** +44(0)1204 532544  
**Fax:** +44(0)1204 522285  
**Email:** [Sales@compactinstruments.co.uk](mailto:Sales@compactinstruments.co.uk)